



ELSEVIER

Available online at [www.sciencedirect.com](http://www.sciencedirect.com)

SCIENCE @ DIRECT®

Nuclear Instruments and Methods in Physics Research A 533 (2004) 612

**NUCLEAR  
INSTRUMENTS  
& METHODS  
IN PHYSICS  
RESEARCH**  
Section A

[www.elsevier.com/locate/nima](http://www.elsevier.com/locate/nima)

Erratum

## Erratum to “Dead time and pileup in pulsed parametric X-ray spectroscopy”

[Nucl. Instr. and Meth. A 524 (2004) 287–294]<sup>☆</sup>

Yaron Danon<sup>\*</sup>, Bryndol Sones, Robert Block

*Department of Mechanical Aerospace and Nuclear Engineering, Rensselaer Polytechnic Institute, NES Bldg,  
Tibbits Ave. 110 8th St., Troy, NY 12180, USA*

Received 23 June 2004

Available online 25 July 2004

---

There is a typographical error in Eq. (12); it is missing a factor of  $f$  in the second term of the right hand side. The correct equation should be

$$n_2 = f \frac{r_2}{f - r_t} - \frac{f}{2} \left( \frac{r_1}{f - r_t} \right)^2 \quad (12)$$

---

<sup>☆</sup> DOI of original articles [10.1016/j.nima.2003.12.034](https://doi.org/10.1016/j.nima.2003.12.034)

<sup>\*</sup> Corresponding author. Tel.: +1-518-276-4008; fax: +1-518-276-4832.  
E-mail address: [danonv@rpi.edu](mailto:danonv@rpi.edu) (Y. Danon).